



Mohammad Waseem Kagzi, M.D.
Diplomat American Board of Sleep Medicine
Diplomat American Board of Internal Medicine

Saquib M. Ahmed, M.D.
Diplomat American Board of Family Medicine

6440 Grand Avenue Ste. 203
Gurnee, IL 60031

Ph: 847-855-9700
Fax: 847-855-8990

Cleaning Water Chamber - For Respironic Remstar Machines

1. Turn the device off and allow approximately 15 minutes for the heater plate and water to cool.
2. Disconnect the tubing from the water chamber. Press down on the water chamber and slide it out of the humidifier platform. Empty any remaining water.
3. Gently remove the base of the chamber with your hands, be careful of the rubber seal.
4. Remove the baffle (piece inside the chamber)(Respironic Water Chambers Only)
5. Use Baby Shampoo and water to wash and rinse them thoroughly.
6. Reassemble and fill the water chamber. Inspect the water chamber for any leaks or damage. Replace the water chamber if any damage is present.

HUMIDIFIER PLATFORM - Respironic & Resmed Machines

WARNING! To avoid electrical shock, disconnect the power cord and power jumper cord before cleaning the humidifier platform. DO NOT immerse the humidifier platform or power jumper cord in any fluids.

1. Clean the humidifier platform by wiping it with a damp cloth. Allow the platform to air dry before reconnecting the power cord.
2. Inspect the humidifier platform and power jumper cord for any damage and replace it if necessary.

Resmed S8 Elite & S8 Auto Machine

1. Open the front latch
2. Remove the water chamber
3. Using the yellow lid provided with the machine, thoroughly rinse the water chamber with warm water and soap. Air dry the device and return it back to its original place
4. Please ensure the lid is tightly shut

Resmed VPAP Machine

1. Take the water chamber out by pressing on the latch underneath the front of the chamber.
2. Open the front lid and wash the inside thoroughly
3. Air dry the device and secure it back into place

SERVICE

1. The humidifier malfunctions or if the green indicator lights start flashing, contact your home health care provider immediately.
(Mohammad W. Kagzi M.D. @ 847-855-9700)

Never attempt to open the humidifier's enclosed

ADDITIONAL NOTES

1. USE ONLY DISTILLED WATER
2. DO NOT OVER FILL WATER CHAMBER. DAMAGE TO THE THERAPY DEVICE MAY OCCUR.
3. DO NOT ALLOW THE CHAMBER TO SIT FOR ANY LENGTH OF TIME AFTER IT HAS BEEN FILLED WITH WATER. IMMEDIATELY INSTALL THE CHAMBER IN THE HUMIDIFIER (STEP 2) ALLOWING WATER TO SIT IN THE WATER CHAMBER (WHEN IT IS NOT CONNECTED TO THE HUMIDIFIER) MAY CAUSE THE CHAMBER TO SEPARATE FROM THE BOTTOM PLATE AND MAY CAUSE WATER LEAKAGE.
4. DO NOT SPILL WATER ON THE C-PAP DEVICE. THIS WILL CAUSE DAMAGE.